

EU Declaration of Conformity

Manufacturer: ndd Medizintechnik AG
Address: Technoparkstrasse 1
CH-8005 Zürich, Switzerland

Herewith we declare our sole responsibility for the declaration of conformity.

We declare under our sole responsibility that the medical device:

Product name: **Easy on-PC**
Product designation: **Spirometry System**
Product type: **Pulmonary Function Testing Device**
Model number: **2700-3**
Classified as: Class IIa
according to annex IX of directive 93/42/EEC

meets all provisions of the directive 93/42/EEC which apply to it.

Applied standards: See Appendix 1

Authorised Representative: **Johner Medical GmbH**
Office Frankfurt
Speicherstrasse 16
60327 Frankfurt am Main, Germany

ndd Medizintechnik AG follows the procedure related to the EC declaration of conformity set out in Annex II of Directive 93/42/EEC which involves the intervention of the Notified Body:


TÜV SÜD Product Service GmbH, Notified Body 0123
Ridlerstrasse 65, 80339 Munich, Germany

Validity of the Declaration of Conformity corresponds to the validity of the EC Certificate
G1 005204 0002 Rev.01.

This declaration of conformity covers the products that have been released for production from the date of issuance of this Declaration of Conformity onward.



Andreas Senn,
Director Quality, Regulatory
Affairs & Clinical Affairs



Georg Harmoncourt,
CEO

Zurich, 25.May.2021

Appendix 1: List of Applied Standards

Standard	Title of standard
EN 60601-1:2006 / A1:2013	Medical electrical equipment, Part 1: General requirements for basic safety and essential performance
EN 60601-1-2:2015	Medical electrical equipment, Part 1-2: General requirements for safety - Collateral standard: Electromagnetic compatibility- Requirements and tests
EN 60601-1-6:2010 / A1:2013	General Requirements for basic safety and essential performance - Collateral standard: Usability
IEC 60601-1-9:2007 / A1:2013	Medical electrical equipment - Part 1-9: General requirements for basic safety and essential performance - Collateral Standard: Requirements for environmentally conscious design
EN 62366:2008 + A1:2015	Medical devices - Application of usability engineering to medical devices
EN 62304:2006 / A1:2015	Medical device software - Software life-cycle processes
EN ISO 14971:2012	Application of risk management to medical devices
EN ISO 26782:2009 / AC:2009	Anaesthetic and respiratory equipment - Spirometers intended for the measurement of time forced expired volumes in humans
EN ISO 23747:2015	Anaesthetic and respiratory equipment - Peak expiratory flow meters for the assessment of pulmonary function in spontaneously breathing humans
ISO 10993-1:2018	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process
EN ISO 10993-5:2009	Biological evaluation of medical devices; part 5: Tests for in vitro cytotoxicity
ISO 10993-10:2010	Biological evaluation of medical devices - part 10: Tests for irritation and skin sensitization
EN ISO 10993-12:2012	Biological evaluation of medical devices - Part 12: Sample preparation and reference materials
EN ISO 10993-18:2009	Biological evaluation of medical devices; part 18: Chemical characterization of materials
EN ISO 15223-1:2016	Medical devices - Symbols to be used with medical device labels, labelling and information to be supplied - Part 1: General requirements
EN 1041:2008 + A1:2013	Information supplied by the manufacturer of medical devices
EN 22248:1992	Packaging - Complete, filled transport packages - Vertical impact test by dropping
ISO 2206:1987	Packaging - Complete, filled transport packages - Identification of parts when testing

Standard	Title of standard
ISO 18562-1:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications – Part 1: Evaluation and testing within a risk management process
ISO 18562-2:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications – Part 2: Tests for emissions of particulate matter
ISO 18562-3:2017	Biocompatibility evaluation of breathing gas pathways in healthcare applications – Part 3: Tests for emissions of volatile organic compounds (VOCs)
MEDDEV 2.7/1 rev.4	Evaluation of clinical data
2012/19/EU	DIRECTIVE 2012/19/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 4 July 2012 on waste electrical and electronic equipment (WEEE)
2011/65/EU	DIRECTIVE 2011/65/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 8 June 2011 on the restriction of the use of certain hazardous substances in electrical and electronic equipment